## Settings of Fluorescence Filter Cube Turntable

**Blank**: Bright Field (for viewing cells and taking BF images)

O: Slider filter (as usual)

**WIY**: YFP (with the Slider @ position #5) **SWB**: CFP (with the Slider @ position #5)